

11-2605

Needle Seat, Std., PEEK™, 0.17mm ID
Comparable to: Agilent® G1329-87017



Technical Data Sheet

1.0 Description

This needle seat assembly is designed to be compatible with Agilent 1100, 1120, 1200, 1220, 1260, G1313A and G1329A/B HPLC systems. It is comparable to Agilent part number G1329-87017.

2.0 Wetted Materials Specifications

316 SS


PEEK

3.0 Replacement Frequency

Replacement may be indicated by leaks at the needle seat or the accumulation of buffer at the top of the needle seat (indicating a small leak), or a blockage of the seat capillary is determined.

4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.