

CTS-0620

Outlet Check Valve Cartridge
Comparable to: Beckman®
240621



Technical Data Sheet

1.0 Description

This check valve is designed to be compatible with Beckman 100A, 110A, 110B, 112, 114M, 116, 118, 125, 126, 127 and 128 HPLC systems. It is comparable to Beckman part number 240621.

2.0 Materials Specifications

316 SS
PTFE
PCTFE
Ruby
Sapphire

3.0 Replacement Frequency

Need for replacement of the check valves may be indicated by system pressure instability, baseline fluctuations, or irregular mobile phase gradients.

4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.