

CTS-10517

Syringe, 500 μ L

Comparable to: Thermo Scientific®

3301-0100



Technical Data Sheet

1.0 Description

This syringe is designed to be compatible with Thermo Scientific 8800 HPLC systems. It is comparable to Thermo Scientific part number 3301-0100.

2.0 Materials Specifications

316 SS

Borosilicate Glass

PTFE

3.0 Replacement Frequency

The syringe may need to be replaced if the plunger tip becomes worn, if injection reproducibility is poor or if the compression check fails.

4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.