

# CTS-10566

Alliance Head Plunger Seal Kit, 2/pk  
Comparable to: Waters® WAT270938 (2 of  
WAT270789)



## Technical Data Sheet

---

### 1.0 Description

These seals are designed to be compatible with Waters 2690, 2690D, 2695, 2695D, 2790, 2795, 2796 and Alliance HPLC systems. It is comparable to Waters part number WAT271018 and 2 of WAT270160.

### 2.0 Wetted Materials Specifications

UHMWPE  
SS Spring

### 3.0 Replacement Frequency

Beyond routine maintenance it may be necessary to replace the seals more frequently if they are operating in a high salt buffer concentration environment or if the plunger themselves are scratched and worn. Seals may also need replacing if the system fails its leak test.

### 4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



### 5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at [sales@sciencix.com](mailto:sales@sciencix.com) to report the issue. We will assist you with a resolution in a timely fashion.