

CTS-10578

500µL Syringe
Comparable to: Shimadzu®
228-25237-04



Technical Data Sheet

1.0 Description

This syringe is designed to be compatible with Shimadzu SIL-10A, SIL-10Ai and SIL-10AXL autosamplers. It is comparable to Shimadzu part number 228-25237-04.

2.0 Wetted Materials Specifications

316 SS
Borosilicate Glass
PTFE

3.0 Replacement Frequency

The syringe may need to be replaced if the plunger tip becomes worn, if injection reproducibility is poor or if the compression check fails.

4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.