

CTS-11288

Isolation Seal for 0101-0921 Valve
Comparable to: Agilent®
0100-1852



Technical Data Sheet

1.0 Description

This isolation seal is designed to be compatible with Agilent 1100 and 1200 HPLC systems. It is comparable to Agilent part number 0100-1852.

2.0 Materials Specifications

Carbon Filled PTFE
SS Spring

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.