

CTS-A126

Outlet Check Valve Assembly
Comparable to: Hitachi® ANO-0837,
810-1005



Technical Data Sheet

1.0 Description

This check valve is designed to be compatible with Hitachi L-2130, L-7100 and L-7110 HPLC systems. It is comparable to Hitachi part number ANO-0837 and 810-1005.

2.0 Materials Specifications

316 SS
Alumina
PCTFE
Zirconium

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.